


# **AEC-NASA TECH BRIEF**



AEC-NASA Tech Briefs describe innovations resulting from the research and development program of the U.S. AEC or from AEC-NASA interagency efforts. They are issued to encourage commercial application. Tech Briefs are published by NASA and may be purchased, at 15 cents each, from the Clearinghouse for Federal Scientific and Technical Information, Springfield, Virginia 22151.

# Cryogenic Liquid Level Measuring Probe



# The problem:

To design a probe to measure the static and dynamic levels of cryogenic liquids in a hydrogen bubble chamber. The sensitive volume, pressure, and density of the liquid are dependent on the liquid height. The probe must be able to perform at -246°C with a short response time.

#### The solution:

A universal liquid level measuring probe, with a unique frequency discriminator, which provides continuous readings on the levels of nitrogen, hydrogen, or helium to an accuracy of  $\pm 1\%$ . The probe has a

dynamic response time of less than 150  $\mu$ sec. allowing boiling conditions or other turbulence to be observed throughout all the transition stages.

### How it's done:

The coaxial probe, immersed in the liquid, measures the difference in the dielectric constants between the liquid and gas phases of the media. The probe capacity is therefore a function of the height of the liquid within the probe elements. The probe and its associated cable constitute the capacity in the tuned circuit of an oscillator and a change in the liquid level is reflected as an oscillator frequency change.

(continued overleaf)

This document was prepared under the sponsorship of the Atomic Energy Commission and/or the National Aeronautics and Space Administration. Neither the United States Government nor any person acting on behalf of the United States Government assumes any liability resulting from the use of the information contained in this document, or warrants that the use of any information, apparatus, method, or process disclosed in this document may not infringe privately owned rights.

A unique feature of the probe is its frequency discriminator. A voltage controlled oscillator is phase-locked to the probe oscillator frequency. The error voltage required to tune the voltage controlled oscillator to the probe oscillator is then a function of the liquid level.

A percentage readout is used with the probe to provide a more efficient use of the instrument's dynamic range and to reduce operator error. In addition, the system may be used in a data logger to record liquid levels in a number of containers through a scanning operation.

## Notes:

- 1. Results are available for a number of test measurements made with N<sub>2</sub> and He<sup>4</sup>.
- 2. Additional details may be found in *The Review of Scient. Instrms.*, vol. 37, no. 11, 1966, p. 1549–1554.

3. Inquiries concerning this innovation may be directed to:

Office of Industrial Cooperation Argonne National Laboratory 9700 South Cass Avenue Argonne, Illinois 60439 Reference: B68-10291

> Source: J. A. Dinkel, Particle Accelerator Division and C. R. Wegner, High Energy Physics (ARG-10138)

#### Patent status:

Inquiries about obtaining rights for commercial use of this innovation may be made to:

Mr. George H. Lee, Chief Chicago Patent Group U.S. Atomic Energy Commission Chicago Operations Office 9800 South Cass Avenue Argonne, Illinois 60439